

Mechanical Stretching Devices for Joint Stiffness and Contracture: Policy Number: CP.MP.144

Last Review Date: 4/20

APPLICATION STATEMENT

The application of the Clinical Coverage Guideline is subject to the benefit determinations set forth by the Centers for Medicare and Medicaid Services (CMS) National and Local Coverage Determinations and state-specific Medicaid mandates, if any.

DISCLAIMER

The Clinical Coverage Guideline (CCG) is intended to supplement certain standard WellCare benefit plans and aid in administering benefits. Federal and state law, contract language, etc. take precedence over the CCG (e.g., Centers for Medicare and Medicaid Services [CMS] National Coverage Determinations [NCDs], Local Coverage Determinations [LCDs] or other published documents). The terms of a member's particular Benefit Plan, Evidence of Coverage, Certificate of Coverage, etc., may differ significantly from this Coverage Position. For example, a member's benefit plan may contain specific exclusions related to the topic addressed in this CCG. Additionally, CCGs relate exclusively to the administration of health benefit plans and are NOT recommendations for treatment, nor should they be used as treatment guidelines. Providers are responsible for the treatment and recommendations provided to the member. The application of the CCG is subject to the benefit determinations set forth by the Centers for Medicare and Medicaid Services (CMS) National and Local Coverage Determinations, and any state-specific Medicaid mandates. Links are current at time of approval by the Medical Policy Committee (MPC) and are subject to change. Lines of business are also subject to change without notice and are noted on www.wellcare.com. Guidelines are also available on the site by selecting the Provider tab, then "Tools" and "Clinical Guidelines".

BACKGROUND

A joint contracture is characterized by chronically reduced ROM secondary to structural changes in non-bony tissues, including muscle, tendons, ligaments, and skin. Prolonged immobilization of joints following surgery or trauma is the most common cause of joint contractures. A number of different modalities are used to treat or prevent joint contractures.

Mechanical stretching devices have been investigated for the treatment of joint contractures. The use of these devices is based on the theory that passive motion early in the healing process can promote movement of the synovial fluid, and thus promote lubrication of the joint; stimulate the healing of articular tissues; prevent adhesions and joint stiffness; and reduce edema without interfering with the healing of incisions or wounds over the moving joint.

Several types of devices exist, including low-load prolonged duration stretch devices (LLPS) (also referred to as dynamic splinting), static progressive stretch devices (SPS), and patient-actuated serial stretch (PASS) (also known as patient-directed serial stretch) devices.

- LLPS devices permit resisted active and passive motion (elastic traction) within a limited range. LLPS devices maintain a set level of tension by means of incorporated springs.
- SPS devices hold the joint in a set position but allow for manual modification of the joint angle and may allow for active motion without resistance (inelastic traction). This type of device itself does not exert a stress on the tissue unless the joint angle is set at the maximum ROM.
- PASS devices permit resisted active and passive motion within a limited range utilizing pneumatic or hydraulic systems that can be adjusted by the patient. The extensionaters use pneumatic systems while the flexionaters

use hydraulic systems. These devices require custom fitting.

Mechanical stretching devices are commonly used in the post-operative period, following an injury or when addressing joint stiffness in the knee, ankle, toe, shoulder, elbow, wrist, or finger. Peer reviewed studies investigating mechanical stretching devices are limited. The best evidence is available in studies evaluating LLPS when used at the knee, elbow, wrist, and following extensor tendon injuries of the finger and for SPS when used at the elbow.

Several authors have looked at the implementation of dynamic splinting at the finger following an extensor tendon repair. Results from a small, prospective, randomized trial comparing dynamic splinting to static splinting suggest that dynamic splinting of complex lacerations of the extensor tendons in zones V-VII provides improved functional outcomes at 4 and 12 weeks and 6 months when compared with static splinting.¹ Another small, prospective, randomized, controlled study comparing postoperative dynamic- versus static- splinting outcomes of patients following extensor tendon repair reported dynamic splinting of simple, complete lacerations of the extensor tendons in zones V and VI. Dynamic splinting provided improved functional outcomes at 4, 6, and 8 weeks but not by 6 months when compared with static splinting.²

Dynamic splinting and static progressive stretch devices have both been applied at the elbow in isolation and in comparison to one another. Gallucci and colleagues (2004) looked at a sample of 30 patients who were at least 78 days after surgery or trauma who had a functional arc of movement of less than 100 degrees at the elbow. They found that 2/3 of patients were able to achieve at least a 100 degree arc and therefore, improved function after using a dynamic splint for 75 days.³ In a randomized controlled pilot study of 30 patients, Lai and colleagues (2009) found significant improvements in ROM when dynamic splinting was added to the control treatment of botulinum toxin type-A and occupational therapy treatment.⁴ Bhat and colleagues (2010) and Gelinas and colleagues (2000) found similar benefit to SPS at the elbow.^{5,6} In both cases, SPS was introduced to the patient approximately 4.5 to 5 months after injury or surgery and once improvements from therapy were stagnant. A functional ROM or arc of movement was achieved in 19 out of 30 patients and 11 out of 22 patients respectively.^{5,6} Doornberg and colleagues (2006) also demonstrated improvements with ROM overall after SPS intervention but noted that early splinting after the initial injury rather than after elbow encapsectomy yielded greater results.⁷ Lindenhovius and colleagues (2012) performed a prospective randomized controlled trial looking at the benefit of dynamic splinting versus SPS in improving range of motion and function as measured by the Disabilities of the Arm, Shoulder, and Hand (DASH).⁸ No significant difference was found between the two groups prior to treatment or after 3, 6 or 12 month follow-ups. Veltman and colleagues (2015) completed a systematic review on the topic that included the results from 232 patients with a similar outcome showing that each device was beneficial but that one was not more effective than the other.⁹

At the knee and wrist, dynamic splinting has been identified as beneficial when further progression of range of motion is needed after surgery or an injury. Pace and colleagues (2018) performed a Level IV retrospective study, looking at the implementation of dynamic splinting following knee surgery in 74 adolescents and children who had ROM deficits in flexion, extension, or both directions.¹⁰ 84% of the patients experienced a significant increase in ROM and 58% were able to avoid further surgical intervention. Willis and colleagues (2016) looked at the treatment of carpal tunnel syndrome using dynamic splinting at the wrist.¹¹ They performed a randomized control trial where the experimental group was provided with dynamic splinting in addition to anti-inflammatories and a stretching program. Those patients who received dynamic splinting in addition to the other treatments had a significant decline in the need for surgical intervention after conservative management was complete. Similarly, Glasgow and colleagues (2011) and Shah and colleagues (2002) looked at the effect of dynamic splinting at the hand and forearm respectively and demonstrated improvements in range of motion after injury in both areas.^{12,13}

Although limited, high-level evidence still exists to address the efficacy of LLPS and SPS interventions, a current review of the literature supports the medical necessity of the current clinical policy. A variety of randomized control trials, observational studies, case series, and medical community acceptance confirms the benefits of dynamic LLPS devices at the knee, elbow, wrist, and fingers and SPS devices at the elbow when used to relieve persistent

joint stiffness that can occur after injury or surgery.

While additional evidence is emerging, there is insufficient evidence in the published peer-reviewed literature to support the use of dynamic LLPS at other joints to include the foot, ankle, and shoulder or SPS devices at any joint other than the elbow. There is insufficient evidence in the published medical literature to demonstrate the safety, efficacy, and long-term outcomes on the use of patient-actuated serial stretch (PASS) devices.

POSITION STATEMENT

Mechanical stretching devices are used for the prevention and treatment of joint contractures of the extremities, with the goal to maintain or restore range of motion (ROM) to the joint. A variety of mechanical stretching devices are available for extension or flexion of the shoulder, elbow, wrist, fingers, knee, ankle, and toes. These devices are generally used as adjunct treatment to physical therapy and/or exercise.

Policy/Criteria

- I. Low-load prolonged-duration stretch (LLPS) device/dynamic stretch device is **medically necessary** for the knee, elbow, wrist or finger when meeting both of the following:
 - A. Meets one of the following indications:
 - 1. In addition to physical therapy in the subacute injury or post-operative period (≥ 3 weeks and ≤ 4 months after injury or operation) in members with signs and symptoms of persistent joint stiffness or contracture;
 - 2. In the subacute injury or post-operative period (≥ 3 weeks and ≤ 4 months after injury or operation) and both of the following:
 - a. Limited range of motion poses a meaningful functional limitation as judged by the physician;
 - b. Has not responded to other therapy (including physical therapy);
 - 3. In the acute post-operative period for members who have undergone additional surgery to improve the range of motion of the previously affected joint;
 - B. Request is for one of the following:
 - 1. An initial four weeks;
 - 2. A subsequent four week period, and improvement was noted upon reevaluation after the prior four week period.
- II. LLPS is considered **experimental/investigational** for any other indication.
- III. Bi-directional static progressive stretch (SPS) devices are considered **experimental/investigational**.
- IV. Patient-actuated serial stretch (PASS) devices are considered **experimental/investigational**.

CODING

HCPCS codes that support coverage criteria

HCPCS Codes	Description
E1800	Dynamic adjustable elbow extension/flexion device, includes soft interface material
E1802	Dynamic adjustable forearm pronation/supination device, includes soft interface
E1805	Dynamic adjustable wrist extension/flexion device, includes soft interface material
E1810	Dynamic adjustable knee extension/flexion device, includes soft interface material

E1812	Dynamic knee, extension/flexion device with active resistance control
E1825	Dynamic adjustable finger extension/flexion device, includes soft interface material

HCPCS codes that do not support coverage criteria

HCPCS Codes	Description
E1399	Durable medical equipment, miscellaneous
E1801	Static progressive stretch elbow device, extension and/or flexion, with or without range of motion adjustment, includes all components and accessories
E1806	Static progressive stretch wrist device, flexion and/or extension, with or without range of motion adjustment, includes all components and accessories
E1811	Static progressive stretch knee device, extension and/or flexion, with or without range of motion adjustment, includes all components and accessories
E1815	Dynamic adjustable ankle extension/flexion device, includes soft interface material
E1816	Static progressive stretch ankle device, flexion and/or extension, with or without range of motion adjustment, includes all components and accessories
E1818	Static progressive stretch forearm pronation/supination device, with or without range of motion adjustment, includes all components and accessories
E1830	Dynamic adjustable toe extension/flexion device, includes soft interface material
E1831	Static progressive stretch toe device, extension and/or flexion, with or without range of motion adjustment, includes all components and accessories
E1840	Dynamic adjustable shoulder flexion/abduction/rotation device, includes soft interface material
E1841	Static progressive stretch shoulder device, with or without range of motion adjustment, includes all components and accessories

ICD-10-CM Diagnosis Codes that Support Coverage Criteria

ICD-10-CM Code	Description
M24.521 - M24.529	Contracture, elbow
M24.531 - M24.539	Contracture, wrist
M24.541 - M24.549	Contracture, hand

M24.561 - M24.569	Contracture, knee
M25.621 - M25.629	Stiffness of elbow, not elsewhere classified
M25.631 - M25.639	Stiffness of wrist, not elsewhere classified
M25.641 - M25.649	Stiffness of hand, not elsewhere classified
M25.661 - M25.669	Stiffness of knee, not elsewhere classified

Coding information is provided for informational purposes only. The inclusion or omission of a CPT, HCPCS, or ICD-10 code does not imply member coverage or provider reimbursement. Consult the member's benefits that are in place at time of service to determine coverage (or non-coverage) as well as applicable federal / state laws.

REFERENCES

1. KitisA, Ozcan RH, Bagdatli D, et al. Comparison of static and dynamic splinting regimens for extensor tendon repairs in zones V to VII. *Plast Surg Hand Surg.* 2012 Sep;46(3-4):267-71
2. Mowlavi A, Burns M, Brown RE. Dynamic versus static splinting of simple zone V and zone VI extensor tendon repairs: a prospective, randomized, controlled study. *Plast Reconstr Surg.* 2005 Feb;115(2):482-7
3. Gallucci GL, Boretto JG, Davalos MA, Alfie VA, Donndorff A, De CP (2014) The use of dynamic orthoses in the treatment of the stiff elbow. *Eur J Orthop Surg Traumatol* 24:1395–1400
4. Lai JM, Francisco GE, Willis FB. Dynamic splinting after treatment with botulinum toxin type-A: a randomized controlled pilot study. *Adv Ther.* 2009 Feb;26(2):241-8.
5. Bhat AK, Bhaskaranand K, Nair SG (2010) Static progressive stretching using a tumbuckle orthosis for elbow stiffness: a prospective study. *J Orthop Surg (Hong Kong)* 18:76–79
6. Gelinas JJ, Faber KJ, Patterson SD, King GJ (2000) The effectiveness of tumbuckle splinting for elbow contractures. *J Bone Joint Surg Br* 82:74–78
7. Doomberg JN, Ring D, Jupiter JB. Static progressive splinting for posttraumatic elbow stiffness. *J Orthop Trauma.* 2006 Jul;20(6):400-4
8. Lindenhovius AL, Doomberg JN, Brouwer KM, Jupiter JB, Mudgal CS, Ring D. A prospective randomized controlled trial of dynamic versus static progressive elbow splinting for posttraumatic elbow stiffness. *J Bone Joint Surg Am.* 2012 Apr 18;94(8):694-700.
9. Veltman ES, Doomberg JN, Eygendaal D, van den Bekerom MP. Static progressive versus dynamic splinting for posttraumatic elbow stiffness: a systematic review of 232 patients. *Arch Orthop Trauma Surg.* 2015 May;135(5):613-7. Pace JL, Nasreddine AY, Simoni M, et al. Dynamic splinting in children and adolescents with stiffness after knee surgery. *J Pediatr Orthop.* 2018 Jan;38(1):38-43.
10. Willis FB, Fowler B. Longitudinal Outcomes Following a Randomized Controlled Trial of Dynamic Splint Stretching for Carpal Tunnel Syndrome. *Hand (N Y).* 2016 Sep;11(3):290-294.
11. Glasgow C, Tooth LR, Fleming J, Peters S. Dynamic splinting for the stiff hand after trauma: predictors of contracture resolution. *J Hand Ther.* 2011;24(3):195-206.
12. Shah MA, Lopez JK, Escalante AS, Green DP. Dynamic splinting of forearm rotational contracture after distal radius fracture. *J Hand Surg Am.* 2002 May;27(3):456-63.
13. Hayes Medical Technology Directory. Mechanical Stretching Device for the Treatment of Joint Contractures of the Extremities. May 9 2018. Update Jun 2019
14. Sameem M, Wood T, Ignacy T, et al. A systematic review of rehabilitation protocols after surgical repair of the extensor tendons in zones V-VIII of the hand. *J Hand Ther.* 2011 Oct-Dec;24(4):365-72
15. Neuhaus V, Wong G, Russo KE, Mudgal CS. Dynamic splinting with early motion following zone IV/V and T1 to TIII extensor tendon repairs. *J Hand Surg Am.* 2012 May;37(5):933-7.
16. Chester DL, Beale S, Beveridge L, Nancarrow JD, Tittley OG. A prospective, controlled, randomized trial comparing early active extension with passive extension using a dynamic splint in the rehabilitation of repaired extensor tendons. *J Hand Surg Br.* 2002;27(3):283-288.
17. Giessler GA, Przybilski M, Germann G, Sauerbier M, Megerle K. Early free active versus dynamic extension splinting after extensor indicis proprius tendon transfer to restore thumb extension: a prospective randomized study. *J Hand Surg Am.* 2008;33(6):864-868
18. Larson D, Jerosch-Herold C. Clinical effectiveness of post-operative splinting after surgical release of Dupuytren's contracture: a systematic review. *BMC Musculoskelet Disord.* 2008 Jul 21;9:104. doi: 10.1186/1471-2474-9-104.
19. Khandwala AR, Webb J, Harris SB, et al. A comparison of dynamic extension splinting and controlled active mobilization of complete divisions of extensor tendons in zones 5 and 6. *J Hand Surg Br.* 2000 Apr;25(2):140-6.
20. Walsh MT, Rinehimer W, Muntzer E et al. Early controlled motion with dynamic splinting versus static splinting for zones III and IV extensor tendon lacerations: a preliminary report. *J Hand Ther.* 1994 Oct-Dec;7(4):232-6.
21. Saldana MJ, Choban S, Westerbeck P, Schacherer TG. Results of acute zone III extensor tendon injuries treated with dynamic extension splinting. *J Hand Surg Am.* 1991 Nov;16(6):1145-50.

22. Griffin M, Hindocha S, Jordan D, et al. Management of Extensor Tendon Injuries. Open Orthop J. 2012; 6: 36–42.
23. Furia JP, Willis FB, Shanumgam R, Curran SA. Systematic review of contracture reduction in the lower extremity with dynamic splinting. Adv Ther. 2013;30(8):763-770.
24. Jongs RA, Harvey LA, Gwinn T, et al. Dynamic splints do not reduce contracture following distal radial fracture: a randomised controlled trial. J Physiother. 2012;58 (3):173-180.

REVIEWS, REVISIONS, and APPROVALS

	Date	Approval Date
Policy developed	04/17	04/17
References reviewed and updated. Codes updated.	03/18	03/18
Removed the following codes from being not medically necessary: E1800, E1801, E1802, E1805, E1810, E1812. Clarified in policy/criteria the joints for which devices are not medically necessary.	03/19	04/19
Added code E1399 as not medically necessary	06/19	
Adapted criteria from WellCare's Dynamic Stretching Devices for Treatment of Joint Stiffness and Contracture HS164. For LPSS, added knee, elbow, and wrist injuries as medically necessary indications. Specified that criteria I.A-I.D be met for LPSS. Removed indication of members unable to benefit from standard physical therapy modalities because of inability to exercise, from original HS164 criteria. Changed the not medically necessary statements regarding LPSS for other indications, PASS and SPS devices to experimental/investigational. Added the following HCPCS codes as supporting coverage criteria: E1800, E1802, E1805, E1810, E1812. Removed HCPCS table of codes not supporting medical necessity. Replaced existing ICD-10 codes with the following: M24.521 - M24.529, M24.531 - M24.539, M24.541 - M24.549, M24.561 - M24.569, M25.621 - M25.629, M25.631 - M25.639, M25.641 - M25.649, M25.661 - M25.669.	4/20	04/20